

Certificate of Conformity

Number CoC-1500734-01 Project number 15200734 Page 1 of 3

Issued by : NMi Certin B.V.

Hugo de Grootplein 1 314 EG Dordrecht The Netherlands

Applicant + + + + +: Schneider Electric dba Power Measurement Ltd.

2195 Keating Cross Road Saanichton, BC V8M 2A5

Canada

Submitted : A meter embedding IEC 61000-4-30 class S Power Quality functions

Manufacturer : Schneider Electric Type : PowerLogic PM8000

Characteristics : See page 2 and further

In accordance with : IEC 61000-4-30 Ed. 2 (2008)

"Electromagnetic Compatibility (EMC) – Part 4-30: Testing and measurement techniques – Power quality measurement methods"

IEC 62586-2 Ed. 1 (2013)

"Power quality measurement in power supply systems - Part 2: Functional

tests and uncertainty requirements'

Measurement class : IEC 61000-4-30 class S

The undersigned declares that the described product is tested according to the above mentioned standard and meet their requirements, based on a non-recurrent examination. The appertaining test data is presented in type evaluation report number NMi-15200734-01 granted by NMi Certin B.V.

NMi Certin B.V. 23 December 2016

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 (0)78 633 23 20 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document only is permitted.





Certificate of Conformity

Number CoC-15200734-01 Project number 15200734 Page 2 of 3

IEC 61000-4-30 Power Quality functions tested

The following IEC 61000-4-30 measurement methods have been tested:

Table 1 IEC 61000-4-30 Power Quality functions tested

IEC 62586-2 Clause	Parameter + + + + + + + + + + + + + + + + + + +	IEC 61000-4-30 class	Comments
+ 7.1 + +	Power frequency + + + + +	+ + +5 + + +	50 and 60 Hz
7.2	Magnitude of the supply voltage	S	+ + + + + + + + + +
7.3	Flicker	+ + + + + +	Not implemented
7.4	Supply voltage interruptions, dips and swells	+ + + ₅ + + + + + + + + +	+ + + + + + + + + + + + + + + + + + + +
+ 7.5+ +	Supply voltage unbalance + + +	+ + +5 + + +	+ + + + + + + + + +
7.6	Voltage harmonics	+ + +5 + + +	+ + + + + + + + +
7.7	Voltage inter-harmonics	+ + + + + +	
+ 7.8+ +	Mains signalling voltages on the voltage supply	+ + ++ + +	Not implemented
7.9 +	Measurement of underdeviation and overdeviation parameters	+ + ++ + +	Not implemented
7.10	Flagging	S	
+ 7.11+ +	Clock uncertainty testing	+ + + + + + +	+ + + + + + + + +
7.12	Variation of external influence quantities	+ + + + + + + + + + + + + + + + + + +	Temperature: -25°C +70°C Power supply: 90 – 415 VAC 120 – 300 VDC
8 + +	Measurement uncertainty and operating uncertainty	+ + + _S + + +	+ + + + + + + + + + + + + + + + + + + +
A: compliance w			: Not implemented



Certificate of Conformity

Number CoC-15200734-01 Project number 15200734 Page 3 of 3

Characteristics of the measuring instrument

In Table 2 the general characteristics of the measuring instrument are presented.

Table 2 General characteristics

U _{din} + + + + + + + + +	230 V _{LN}		
fnom + + + + + + + + +	50 Hz and 60 Hz		
Temperature + + + + + +	Rated range of operation: -25°C to +70°C		
Power supply range	90 – 415 VAC (+/- 10%), 50/60 Hz 120 – 300 VDC (+/- 10%), 45 VA / 20 W		
Software version	001.000.057		
Hardware version + + +	METSEPM8240 DIN96 panel mount meter METSEPM8243 DIN rail mount meter		
Environmental application	Fixed (F), Indoor (I)		