

Certificate of Conformity

 (\bullet)

Number CoC-2274923-02 Project number 2274923 Page 1 of 3

Issued by	: NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands	
Applicant	: Schneider Electric dba Power Measurement Ltd. 2195 Keating Cross Road Saanichton, BC V8M 2A5 Canada	
Submitted	: A meter embedding IEC 61000-4-30 class A Power Quality functions	
	Manufacturer : Schneider Electric Type : PowerLogic ION7400	
Characteristics	: See page 2 and further	
In accordance with	 : IEC 61000-4-30 Ed. 3 (2015) "Electromagnetic Compatibility (EMC) – Part 4-30: Testing and measurement techniques – Power quality measurement methods" IEC 62586-2 Ed. 2 (2017) "Power quality measurement in power supply systems - Part 2: Functional tests and uncertainty requirements" 	
Measurement class	: IEC 61000-4-30 class S	

The undersigned declares that the described product is tested according to the above mentioned standard and meet their requirements, based on a non-recurrent examination. The appertaining test data is presented in type evaluation report number NMi-15200734-01 and NMi-2274923-01, granted by NMi Certin B.V.

NMi Certin B.V. 8 March 2019

C. Oosterman Head Certification Board

NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands T +31 88 636 2332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document only is permitted.







Certificate of Conformity

Number CoC-2274923-02 Project number 2274923 Page 2 of 3

> VAC VDC

IEC 61000-4-30 Power Quality functions tested

The following IEC 61000-4-30 measurement methods have been tested

IEC 62586-2 Clause	Parameter	IEC 61000-4-30 class	Comments
7.1	Power frequency	S	50 and 60 Hz
7.2	Magnitude of supply voltage	S	
7.3	Flicker	S	Class F3 230V, 50 Hz 120V, 60 Hz
7.4	Supply voltage interruptions, dips and swells	s	50 and 60 Hz
7.5	Supply voltage unbalance	S	
7.6	Voltage harmonics	S	
7.7	Voltage interharmonics		
7.8	Mains signalling voltages on the voltage supply	S	Method 2
7.9	Measurement of underdeviation and overdeviation parameters		Not applicable for class S
7.10	Flagging	S	
7.11	Clock uncertainty testing	S	
7.12	Variation of external influence quantities	S	Temperature: -25°C +70°C Power supply: 90 – 415 VAC 110 – 415 VDC
7.13	Rapid Voltage Changes (RVC)		
7.14	Magnitude of current	S	
7.15	Harmonic current	S -	
7.16	Interharmonic currents		
7.17	Current unbalance	S	
8	Calculation of measurement uncertainty and operating uncertainty	S	

: compliance with class A Α

S : compliance with class S : Not implemented

The tests are performed in accordance with IEC 62586-2 edition 2 (2017).



Certificate of Conformity



Number CoC-2274923-02 Project number 2274923 Page 3 of 3

Characteristics of the measuring instrument In the following table, the general characteristics of the measuring instrument are presented.

General characteristics

eneral characteristics				
Model	ION7400			
U _{din}	230 V _{LN}			
I _{nom}	1 A, 5 A			
$f_{\sf nom}$	50 Hz and 60 Hz			
Temperature	Rated range of operation: -25°C to +70°C			
Power supply range	90 – 415 VAC (+/- 10%), 50/60 Hz 110 – 415 VDC (+/- 10%)			
Software version	002.000.xxx			
Hardware version	METSEION74XX			
Environmental application Fixed (F), Indoor (I)				



